


# NASA TECH BRIEF



NASA Tech Briefs are issued by the Technology Utilization Division to summarize specific technical innovations derived from the space program. Copies are available to the public from the Clearinghouse for Federal Scientific and Technical Information, Springfield, Virginia, 22151.

## Fluid-Pressure Measurement Apparatus Uses Short-Length Manometer Tubes



**The problem:** Measuring high fluid pressures by means of manometers. In situations where high pressures occur, an inordinately long column of liquid and tube would be required in a conventional manometer. Liquid blowout could also occur from the open end of such a manometer during pressure surges.

**The solution:** Use of a system of short-length U-tube manometers in conjunction with a reference pressure which is divided into proportional parts.

**How it's done:** A reference pressure  $P_R$  divided into proportions by (Kendall) ratio boosters is introduced into a U-tube system. The illustration shows three separate U-tubes, with the indicated proportions of the reference pressure applied to the different branches. With this proportioning of the pressures, the difference in the heights,  $h_1$ ,  $h_2$ , and  $h_3$ , of the liquid levels in each U-tube is approximately the same ( $h_1 \approx h_2 \approx h_3 \approx 1/3 P_R$ ), and the reference pressure can therefore be determined by arithmetically adding the

(continued overleaf)

three level differentials. The reference pressure is also introduced into one branch of a common-well manometer, and the pressure to be measured,  $P_X$ , is applied to another branch of the common-well manometer. It is easy to show that for this configuration,  $P_X = P_R + H$ , or  $P_X = H + h_1 + h_2 + h_3$ , where  $H$  is the difference in the heights of the liquid levels in the two branches of the manometer in the common well. Since the last branch of the U-tube is open to the atmosphere,  $P_X$  is the gage pressure (in terms of the height of the liquid used in the manometer system).

**Notes:**

1. The reference pressure may be supplied from any convenient source within the pressure system. This pressure need not be constant.
2. Inquiries concerning this innovation may be directed to:

Technology Utilization Officer  
Lewis Research Center  
21000 Brookpark Road  
Cleveland, Ohio, 44135  
Reference: B65-10027

**Patent status:** NASA encourages commercial use of this innovation. No patent action is contemplated.

Source: B.I. Sather  
(Lewis-28)